

**EC Declaration of Conformity**

DOCM02ver3

Manufacturer: NORECS AS  
Address: Gaustadalleen 21, NO-0349 Oslo, Norway

We NORECS AS declare under our sole responsibility that the Product

Model: ProboStat  
Type: A-1, A-2, A-3, A-4, A-5, A-6  
Description: Passive sample holder for scientific electrical measurements

Is in conformity with the provisions of the European Council Pressure Equipment Directive (PED) 97/23/EC of 29 May 1997 which apply.

Compliance with this directive is under Sound Engineering Practice (SEP).

Assumptions of conformity are based on the following: The Product may be used with fluids (gases of Type 1 or Type 2, ref. Article 9) of pressures above 0.5 bar over atmospheric pressure. The maximum allowed pressure is 15 bar. The maximum volume is 1 litre. The parts are assembled with a heated vessel (ref. Article 3 paragraph 1.1 and 1.2) and tubing of DN 6 (ref. Article 3 paragraph 1.3) with less than 1 L total volume. According to Annex II, Tables 1, 2, 3, 4, and 5, the Product falls below the lower limits of Category I. According to Article 3 paragraph 3 it therefore shall not and must not be EC certified under PED.

The Product generates voltages less than 2 V DC. It is to be used with externally applied voltages for measurements or internal heating of less than 40 V DC or AC (p-p).

Declaration of Conformity valid since: 27.09.2017



A handwritten signature in blue ink that reads "Ingvild Lorentzen".

Managing Director  
NORECS AS

Signed: Ingvild Lorentzen  
Date: 02.10.2017